

## Implant Compatibility List for NeoBiotech Implant Removal Kit

ПРИМЕЧАНИЕ: Для других производителей, не указанных в списке, врачу рекомендуется выполнить "пробную установку", выбрав соответствующий аналог или имплантат для нарезания резьбы, при этом последовательно начиная с самого широкого диаметра винта FRS M2.5, затем в порядке убывания - следующий доступный диаметр. Это обеспечит совместимость.

Corporation	Implant System	Implant Ø Diameter	Screw Size to use	Solution Kit Compatibility	Product Name
Nobel Biocare®	Branemark System® MK III Groovy NP	3.3mm	M1.6	YES	FRS 16
	Branemark System MK III Groovy RP	3.75mm	M2.0	YES	FRS 20
	Branemark System MK III Groovy RP	4.0mm	M2.0	YES	FRS 20
	Branemark System MK III Groovy WP	5.0mm	M2.5	YES	FRS 25
	NobelSpeedy® System MK III Groovy NP	3.3mm	M1.6	YES	FRS 16
	NobelSpeedy System MK III Groovy RP	4.0mm	M2.0	YES	FRS 20
	NobelSpeedy System MK III Groovy WP	5.0mm	M2.0	YES	FRS 20
	NobelSpeedy System MK III Groovy WP	6.0mm	M2.5	YES	FRS 25
	NobelReplace® Straight Groovy NP	3.5mm	M1.8	YES	FRS 18
	NobelReplace Straight Groovy RP	4.0mm	M2.0	YES	FRS 20
	NobelReplace Straight Groovy WP	5.0mm	M2.0	YES	FRS 20
	NobelSpeedy Replace NP	3.5mm	M1.8	YES	FRS 18
	NobelSpeedy Replace RP	4.0mm	M2.0	YES	FRS 20
	NobelSpeedy Replace WP	5.0mm	M2.0	YES	FRS 20
	NobelSpeedy Replace 6.0	6.0mm	M2.0	YES	FRS 20
	NobelReplace Tapered NP	3.5mm	M1.8	YES	FRS 18
	NobelReplace Tapered RP	4.3mm	M2.0	YES	FRS 20
	NobelReplace Tapered WP	5.0mm	M2.0	YES	FRS 20
	NobelReplace Tapered	6.0mm	M2.0	YES	FRS 20
	NobelActive®	3.5mm	M1.54	SPECIAL ORDER	FRS 154
Straumann®	Standard	3.3 RN	M2.0	YES	FRS 20
	Standard	4.1 RN	M2.0	YES	FRS 20
	Standard	4.8 RN	M2.0	YES	FRS 20
	Standard	4.8 WN	M2.0	YES	FRS 20
	Standard Plus	3.3 NN	M1.6 or M1.8	Verify	FRS 16 / FRS 18
	Standard Plus	3.3 RN	M2.0	YES	FRS 20
	Standard Plus	4.1 RN	M2.0	YES	FRS 20
	Standard Plus	4.8 RN	M2.0	YES	FRS 20
	Standard Plus	4.8 WN	M2.0	YES	FRS 20
	Tapered Effect	3.3 RN	M2.0	YES	FRS 20
	Tapered Effect	4.1 RN	M2.0	YES	FRS 20
	Tapered Effect	4.8 RN	M2.0	YES	FRS 20
	Bone Level Implant	3.3 NC	M1.6 or M1.8	Verify	FRS 16 / FRS 18
	Bone Level Implant	4.1 NC	M1.6 or M1.8	Verify	FRS 16 / FRS 18
	Bone Level Implant	4.8 NC	M1.6 or M1.8	Verify	FRS 16 / FRS 18
BIOMET 3i	Certain® Internal	3.4	M1.6	YES	FRS 16
	Certain Internal	4.1	M1.6	YES	FRS 16
	Certain Internal	5.0	M1.6	YES	FRS 16

# Implant Compatibility List for NeoBiotech Implant Removal Kit

NOTE: For other Manufacturers not listed, the clinician is advised to complete a "try in" by selecting a corresponding analog or implant to thread into whereby sequentially starting with the widest diameter FRS screw M2.5 followed by, in descending order, the next available diameter. This will ensure compatibility.

Corporation	Implant System	Implant Ø Diameter	Screw Size to use	Solution Kit Compatibility	Product Name
	Certain® Internal	6.0	M1.6	YES	FRS 16
	External	3.4	M2.0	YES	FRS 20
	External	4.1	M2.0	YES	FRS 20
	External	5.0	M2.0	YES	FRS 20
	External	6.0	M2.0	YES	FRS 20
Zimmer® (Screw-Vent®)	Screw Advent Implant	3.3	1-72UNF	YES	FRS 172
	Screw Advent Implant	3.7	1-72UNF	YES	FRS 172
	Screw Advent Implant	4.7	1-72UNF	YES	FRS 172
	Tapered Screw Advent Implant	3.7	1-72UNF	YES	FRS 172
	Tapered Screw Advent Implant	4.7	1-72UNF	YES	FRS 172
	Tapered Screw Advent Implant	6.0	1-72UNF	YES	FRS 172
	Advent® Implant	3.7	1-72UNF	YES	FRS 172
	Advent Implant	4.7	1-72UNF	YES	FRS 172
	Advent Implant	4.7	1-72UNF	YES	FRS 172
Zimmer® (Swissplus®)	Straight Swissplus Implant	4.1	M2.0	YES	FRS 20
	Straight Swissplus Implant	4.8	M2.0	YES	FRS 20
	Tapered Swissplus Implant	3.7	1-72UNF	YES	FRS 172
	Tapered Swissplus Implant	3.7	M2.0	YES	FRS 20
	Tapered Swissplus Implant	4.8	M2.0	YES	FRS 20
Zimmer® (Spline®)	Spline Twist Implants	3.75mmD	1-72UNF	YES	FRS 172
	Spline Twist Implants	5.0mmD	1-72UNF	YES	FRS 172
	Spline Reliance Cylinder Implants	3.25mmD	1-72UNF	Verify	FRS 172
	Spline Reliance Cylinder Implants	4.0mmD	1-72UNF	YES	FRS 172
	Spline Reliance Cylinder Implants	5.0mmD	1-72UNF	YES	FRS 172
Astra Tech®	OsseoSpeed	3.0S	M1.4	NO	FRS 14
	OsseoSpeed	3.5S	M1.6	YES	FRS 16
	OsseoSpeed	4.0S	M1.6	YES	FRS 16
	OsseoSpeed	4.5	M2.0	YES	FRS 20
	OsseoSpeed	5.0	M2.0	YES	FRS 20
	OsseoSpeed	5.0S	M2.0	YES	FRS 20
Keystone Dental (Lifecore)	PrimaConnex® Internal Connection Tapered Implants	3.5	M1.8	YES	FRS 18
	PrimaConnex Internal Connection Tapered Implants	4.1	M1.8	YES	FRS 18
	PrimaConnex Internal Connection Tapered Implants	5.0	M1.8	YES	FRS 18
	PrimaConnex Internal Connection Straight Implants	3.3	M1.8	YES	FRS 18
	PrimaConnex Internal Connection Straight Implants	4.0	M1.8	YES	FRS 18
	PrimaConnex Internal Connection Straight Implants	5.0	M1.8	YES	FRS 18
	RENOVA® RBM Tapered Implants	3.75	M1.8	YES	FRS 18
	RENOVA RBM Tapered Implants	4.5/4.75	M1.8	YES	FRS 18

# Implant Compatibility List for NeoBiotech Implant Removal Kit

NOTE: For other Manufacturers not listed, the clinician is advised to complete a "try in" by selecting a corresponding analog or implant to thread into whereby sequentially starting with the widest diameter FRS screw M2.5 followed by, in descending order, the next available diameter. This will ensure compatibility.

Corporation	Implant System	Implant Ø Diameter	Screw Size to use	Solution Kit Compatibility	Product Name
	RENOVA RBM Straight Implants	3.75	M1.8	YES	FRS 18
	RENOVA RBM Straight Implants	4.5/4.75	M1.8	YES	FRS 18
	RESTORE® Small Diameter	3.3	1-72UNF	YES	FRS 172
	RESTORE Regular Diameter	3.75 & 4.0	M2.0	YES	FRS 20
	RESTORE Small Diameter Wide Diameter	5.0 & 6.0	M2.5	YES	FRS 25
	STAGE-1® system	3.3 (RDS)	M2.0	YES	FRS 20
	STAGE-1 system	4.1 (RDS)	M2.0	YES	FRS 20
	STAGE-1 system	4.8 (RDS)	M2.0	YES	FRS 20
	STAGE-1 system	4.8 (WDS)	M2.0	YES	FRS 20
	STAGE-1 system	5.5 (WDS)	M2.0	YES	FRS 20
	STAGE-1 system	6.3 (WDS)	M2.0	YES	FRS 20
MIS®	MISTRAL Internal Octagon System	3.75	M2.0	YES	FRS 20
	MISTRAL Internal Octagon System	4.1	M2.0	YES	FRS 20
	MISTRAL Internal Octagon System	4.8	M2.0	YES	FRS 20
	Biocom Internal	3.3	1-72UNF	YES	FRS 172
	Biocom Internal	3.75	1-72UNF	YES	FRS 172
	Biocom Internal	4.2	1-72UNF	YES	FRS 172
	Biocom Internal	5.0	1-72UNF	YES	FRS 172
	Biocom Internal	6.0	1-72UNF	YES	FRS 172
	Seven® Internal	3.75	1-72UNF	YES	FRS 172
	Seven Internal	4.2	1-72UNF	YES	FRS 172
	Seven Internal	5.0	1-72UNF	YES	FRS 172
	Seven Internal	6.0	1-72UNF	YES	FRS 172
	Lance Internal	3.75	1-72UNF	YES	FRS 172
	Lance Internal	4.2	1-72UNF	YES	FRS 172
	Lance Internal	5.0	1-72UNF	YES	FRS 172
Dentsply™	Ankylos® internal	3.5	M1.8	YES	FRS 18
	Ankylos internal	4.5	M1.8	YES	FRS 18
	Ankylos internal	5.5	M1.8	YES	FRS 18
	Ankylos internal	7.0	M1.8	YES	FRS 18
	Xive® S internal Implant	3.0	M1.6	YES	FRS 16
	Xive S internal Implant	3.4	M1.6	YES	FRS 16
	Xive S internal Implant	3.8	M1.6	YES	FRS 16
	Xive S internal Implant	4.5	M1.6	YES	FRS 16
	Xive S internal Implant	5.5	M1.6	YES	FRS 16
	Xive TG External Implant	3.4	M1.4	Verify	FRS 14
	Xive TG External Implant	3.8	M1.4	Verify	FRS 14
	Xive TG External Implant	4.5	M1.4	Verify	FRS 14

# Implant Compatibility List for NeoBiotech Implant Removal Kit

NOTE: For other Manufacturers not listed, the clinician is advised to complete a "try in" by selecting a corresponding analog or implant to thread into whereby sequentially starting with the widest diameter FRS screw M2.5 followed by, in descending order, the next available diameter. This will ensure compatibility.

Corporation	Implant System	Implant Ø Diameter	Screw Size to use	Solution Kit Compatibility	Product Name
	Frialit® Internal Implant	3.4	M1.6	YES	FRS 16
	Frialit Internal Implant	3.8	M1.6	YES	FRS 16
	Frialit Internal Implant	4.5	M1.6	YES	FRS 16
	Frialit Internal Implant	5.5	M1.6	YES	FRS 16
	Frialit Internal Implant	6.5	M1.6	YES	FRS 16
CAMLOG®	CAMLOG Root Line	3.8	M1.6	YES	FRS 16
	CAMLOG Root Line	4.3	M1.6	YES	FRS 16
	CAMLOG Root Line	5.0	M2.0	YES	FRS 20
	CAMLOG Root Line	6.0	M2.0	YES	FRS 20
Biohorizons®	Maestro® External implants	3.5	M2.0	Verify	FRS 20
	Maestro External implants	4.0	M2.0	Verify	FRS 20
	Maestro External implants	5.0	M2.0	Verify	FRS 20
	Maestro External implants	6.0	M2.0	Verify	FRS 20
	Biohorizons Internal	3.5	1-72UNF	Verify	FRS 172
	Biohorizons Internal	4.0	1-72UNF	Verify	FRS 172
	Biohorizons Internal	5.7	1-72UNF	Verify	FRS 172
SPI® Implant	SPI Internal Implant	3.5	M1.4	Verify	FRS 14
	SPI Internal Implant	4.5	M1.6	YES	FRS 16
	SPI Internal Implant	5.0	M1.6	YES	FRS 16
	SPI Internal Implant	6.0	M1.6	YES	FRS 16

ICL-IRK - 3/14-REV-3)

Distributed by:



4555 Riverside Drive  
 Palm Beach Gardens, FL 33410  
 1-800-342-5454  
 Outside The U.S.: +1-561-776-6700  
 Fax: +1-561-776-1272  
 www.biomet3i.com

Manufactured by:



e-space # 103, 104-1, 104-2, 105,  
 106, 205, 212, 312, 509, 510, 511,  
 10F, 36, 27 Gil, Digital-ro, Guro-gu,  
 SEOUL, 152-789  
 Rep. of Korea  
 Tel.: 82-2-582-2885  
 Fax: 82-2-582-2883  
 www.neobiotech.co.kr  
 Made in Korea

Branemark System, NobelActive, NobelReplace and NobelSpeedy are registered trademarks of Nobel Biocare. Camlog is a registered trademark of Camlog Biotechnologies AG. Nobel Biocare is a registered trademark of Nobel Biocare Services AG. Straumann is a registered trademark of Straumann Holding AG. Astra Tech, Frialit, Xive and Osseospeed are trademarks of Dentsply. NeoBiotech is a trademark of NEOBIOTECH CO., LTD. Advent, Screw-Vent, SwissPlus, Spine and Zimmer are registered trademarks of Zimmer Dental, Inc. Maestro and BioHorizons are registered trademarks of BioHorizons Implant Systems Inc. PrimaConnex, Renova and Stage-1 are registered trademarks of Keystone Dental, Inc. MIS and Seven are registered trademarks of Mis-Implants Technologies Ltd. SPI is a registered trademark of Thommen Medical. Certain is a registered trademark and Providing Solutions - One Patient At A Time and design are trademarks of BIOMET 3i LLC. BIOMET 3i and design are trademarks of BIOMET, Inc.  
 © 2014 BIOMET 3i LLC. All rights reserved.